

Konformitätserklärung
Declaration of Conformity

Wir

We

B. Braun Melsungen AG
Carl-Braun-Straße 1
34212 Melsungen
Deutschland/Germanyerklären in eigener Verantwortung,
dass das/die Produkt/e**Fixierplatte für Ureofix® 500 N,**
Schlauchfixierklemme für Ureofix®,
U500 N Adapter für Hill-Rom-Betten,
Ureofix® 500 N FixierbandZubehör
(Artikelnummern siehe Anlage I)mit den Anforderungen der folgenden Richtlinie
übereinstimmt/übereinstimmenRichtlinie 93/42/EWG des Rates vom 14. Juni 1993
über Medizinprodukte**Konformitätsbewertungsverfahren**
nach Anhang VII
der oben genannten Richtlinie**Klassifizierung**
gemäß Anhang IX
der oben genannten Richtlinie
Klasse I**Datum der ersten CE-Kennzeichnung**
1994-12

Melsungen, 2008-03-17

B. Braun Melsungen AG

i. V.

J. Heil
Vice President Quality & Environment Managementhereby declare in our own responsibility
that the product/s**Fixation plate for Ureofix® 500 N,**
Tube Fixation Clamp for Ureofix®,
U500 N Adapter for Hill-Rom-Beds,
Ureofix® 500 N Fixation StrapsAccessories
(article numbers see attachment I)


is/are in compliance with the following directive

Council Directive 93/42/EEC of 14th June 1993
concerning Medical Devices**Conformity Assessment Procedure**
according to annex VII
of the Council Directive named above**Classification**
according to annex IX
of the Council Directive named above
Class I**Date of first CE-marking**
1994-12

Melsungen, 2008-03-17

B. Braun Melsungen AG

i. V.

G. Henke
Manager Regulatory Affairs CoE IV-Systems**Effective**

Konformitätserklärung
Declaration of Conformity**Anlage I / Attachment I**

Art.-Nr. / Artikelbezeichnung Art. no.		Article description	Klasse / Class
4411404	Schlauchfixierklemme für Ureofix®	Tube Fixation Clamp for Ureofix®	I
4417755	Ureofix® 500 N Fixierband	Ureofix® 500 N Fixation Straps	I
4417780	Fixierplatte für Ureofix® 500 N	Fixation plate for Ureofix® 500 N	I
4417798	U500 N Adapter für Hill-Rom-Betten	U500 N Adapter for Hill-Rom-Beds	I

Effective